

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ Directional RNA Library Prep Kit for Illumina®
 Catalog Number: E7420L
 Packaging Lot Number: 10125930
 Expiration Date: 11/2022
 Storage Temperature: -20°C
 Specification Version: PS-E7420S/L v1.0

NEBNext® Ultra™ Directional RNA Library Prep Kit for Illumina® Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7431AAVIAL	Nuclease-free Water	10108909	Pass
E7428AAVIAL	NEBNext® USER® Enzyme	10108910	Pass
E7426AAVIAL	NEBNext® Second Strand Synthesis Reaction Buffer (dUTP Mix)	10108905	Pass
E7425AAVIAL	NEBNext® Second Strand Synthesis Enzyme Mix	10108904	Pass
E7424AAVIAL	RNase Inhibitor, Murine	10108903	Pass
E7423AAVIAL	ProtoScript® II Reverse Transcriptase	10108912	Pass
E7422AAVIAL	Random Primers	10108902	Pass
E7421AAVIAL	NEBNext® First Strand Synthesis Reaction Buffer	10108901	Pass
E7373AAVIAL	Blunt/TA Ligase Master Mix	10108908	Pass
E7372AAVIAL	NEBNext® End Repair Reaction Buffer	10108907	Pass
E7371AAVIAL	NEBNext® End Prep Enzyme Mix	10108906	Pass
E6625AAVIAL	NEBNext® Q5® Hot Start HiFi PCR Master Mix	10108913	Pass

Assay Name/Specification	Lot # 10125930
<p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Ultra™ Directional RNA Library Prep Kit for Illumina® and meet the designated specifications.</p>	Pass
<p>Functional Testing (Library Construction, RNA) Each set of reagents is functionally validated and compared to a previous lot through construction of libraries made from commercially available RNA, using the kit's minimum and maximum input requirements. Libraries made from previous and current lots for both input RNA amounts are sequenced together on the same Illumina® flow cell and compared across various metrics including library yield, individual transcript abundance correlations (low vs. high input, old lot vs. new lot), 5'-3'</p>	Pass

Assay Name/Specification	Lot # 10125930
transcript coverage, and fraction of reads mapping to a reference.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Christine Sumner
Production Scientist
02 Dec 2021



Michael Tonello
Packaging Quality Control Inspector
02 Dec 2021