

## New England Biolabs Certificate of Analysis

Product Name: NEBNext<sup>®</sup> Ultra<sup>™</sup> II FS DNA Module  
 Catalog Number: E7810L  
 Packaging Lot Number: 10196355  
 Expiration Date: 05/2024  
 Storage Temperature: -20°C  
 Specification Version: PS-E7810S/L v1.0

NEBNext <sup>®</sup> Ultra <sup>™</sup> II FS DNA Module Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7808AAVIAL	TE Buffer	10187218	Pass
E7807AAVIAL	NEBNext <sup>®</sup> Ultra <sup>™</sup> II FS Reaction Buffer	10187217	Pass
E7806AAVIAL	NEBNext <sup>®</sup> Ultra <sup>™</sup> II FS Enzyme Mix	10187216	Pass

Assay Name/Specification	Lot # 10196355
<p><b>* Individual Product Component Note</b>            Standard Quality Control Tests are performed for each component included in NEBNext<sup>®</sup> Ultra<sup>™</sup> II FS DNA Module and meet the designated specifications.</p> <p><b>Functional Testing (Library Construction, FS DNA)</b>            Each set of reagents is functionally validated and compared to the previous lot through construction of libraries made from commercially available genomic DNA, using the kit's minimum and maximum input requirements. A fragmentation time of 20 minutes was used to generate an insert size of approximately 200 bp. The final average library size is between 270 and 450 bp as determined by an Agilent Bioanalyzer. Libraries made from the previous and current lots for both input DNA amounts are sequenced together on the same Illumina flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.</p>	<p>Pass</p> <p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.

*Christine Sumner*

---

Christine Sumner  
Production Scientist  
08 Jun 2023

*Josh Hersey*

---

Josh Hersey  
Packaging Quality Control Inspector  
08 Jun 2023