

New England Biolabs Certificate of Analysis

Product Name: *FseI*
Catalog Number: *R0588S*
Concentration: *2,000 U/ml*
Unit Definition: *One unit is defined as the amount of enzyme required to digest 1 µg of pBC4 DNA in 1 hour at 37°C in a total reaction volume of 50 µl.*
Packaging Lot Number: *10173228*
Expiration Date: *08/2023*
Storage Temperature: *-80°C*
Storage Conditions: *10 mM Tris-HCl , 100 mM KCl , 1 mM DTT , 0.1 mM EDTA , 0.5 % Tween® 20 , 0.5 % IGEPAL® CA-630 , 50 % Glycerol, (pH 7.4 @ 25°C)*
Specification Version: *PS-R0588S/L v3.0*


FseI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0588SVIAL	FseI	10162512	Pass
B7024AVIAL	Gel Loading Dye, Purple (6X)	10162784	Pass
B6004SVIAL	rCutSmart™ Buffer	10165689	Pass

Assay Name/Specification	Lot # 10173228
Protein Purity Assay (SDS-PAGE) FseI is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 10 Units of FseI incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 10 units of FseI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of pBC4 DNA and a minimum of 10 units of FseI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

Assay Name/Specification	Lot # 10173228
<p>Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of pBC4 DNA with FseI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with FseI.</p>	<p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



YunJie Sun
Production Scientist
30 Aug 2022



Michael Tonello
Packaging Quality Control Inspector
29 Nov 2022