

New England Biolabs Certificate of Analysis

Product Name: PstI-HF[®]
Catalog Number: R3140M
Concentration: 100,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Packaging Lot Number: 10090258
Expiration Date: 10/2022
Storage Temperature: -20°C
Storage Conditions: 250 mM NaCl, 10 mM Tris-HCl, 1 mM DTT, 0.1 mM EDTA, 50 % Glycerol, 0.15 % Triton X-100, 200 µg/ml BSA, (pH 7.4 @ 25°C)
Specification Version: PS-R3140T/M v2.0

PstI-HF [®] Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R3140M VIAL	PstI-HF [®]	10088041	Pass
B7204S VIAL	CutSmart [®] Buffer	10085423	Pass
B7024A VIAL	Gel Loading Dye, Purple (6X)	10084971	Pass

Assay Name/Specification	Lot # 10090258
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart [®] Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 200 units of PstI-HF [®] incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 100-fold over-digestion of Lambda DNA with PstI-HF [®] , >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with PstI-HF [®] .	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart [®] Buffer containing 1 µg of Lambda DNA and a minimum of 200 units of PstI-HF [®] incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Protein Purity Assay (SDS-PAGE) PstI-HF [®] is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Penghua Zhang
Production Scientist
05 Nov 2020



Michael Tonello
Packaging Quality Control Inspector
05 Nov 2020