

New England Biolabs Certificate of Analysis

Product Name: *NruI-HF*
Catalog #: *R3192S/L*
Concentration: *20,000 units/ml*
Lot #: *0021603*
Assay Date: *03/2016*
Expiration Date: *3/2018*
Storage Temp: *-20°C*
Storage Conditions: *50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50 % Glycerol, 200 µg/ml BSA*
Specification Version: *PS-R3192S/L v1.0*
Effective Date: *31 Oct 2014*

Assay Name/Specification (minimum release criteria)	Lot #0021603
Endonuclease Activity (Nicking) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled pUC19 DNA and a minimum of 100 units of NruI-HF incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] <i>E. coli</i> DNA and a minimum of 100 units of NruI-HF incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Functional Test (15 minute Digest) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and 1 µl of NruI-HF incubated for 15 minutes at 37°C results in complete digestion as determined by agarose gel electrophoresis.	Pass
Ligation and Recutting (Terminal Integrity) - After a 20-fold over-digestion of Lambda DNA with NruI-HF, ~50% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with NruI-HF.	Pass
Non-Specific DNase Activity (16 Hour) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 100 units of NruI-HF incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Protein Purity Assay (SDS-PAGE) - NruI-HF is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass



New England Biolabs Certificate of Analysis

* The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.



Authorized by
Derek Robinson
31 Oct 2014



Inspected by
Jianying Luo
08 Mar 2016

