

New England Biolabs Certificate of Analysis


Product Name: *Monarch[®] Plasmid Miniprep Kit*
 Catalog Number: *T1010G*
 Packaging Lot Number: *10208627*
 Expiration Date: *10/2025*
 Storage Temperature: *25°C*
 Specification Version: *PS-T1010G v2.0*

Monarch [®] Plasmid Miniprep Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T1017-21	Monarch [®] Plasmid Miniprep Columns	10208162	Pass
T1016-21	Monarch [®] DNA Elution Buffer	10208158	Pass
T1015-21	Monarch [®] Plasmid Wash Buffer 2	10208153	Pass
T1014-21	Monarch [®] Plasmid Wash Buffer 1	10208146	Pass
T1013-21	Monarch [®] Plasmid Neutralization Buffer (B3)	10208143	Pass
T1012-21	Monarch [®] Plasmid Lysis Buffer (B2)	10208139	Pass
T1011-21	Monarch [®] Plasmid Resuspension Buffer (B1)	10208129	Pass

Assay Name/Specification	Lot # 10208627
Functional Testing (DNA Recovery and Purity) Twenty-four Monarch [®] Plasmid Miniprep Columns are tested with 1.5 µg of 1 Kb Ladder resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	Pass
Functional Testing (Plasmid Purification and Restriction Digestion) pUC19 plasmid is purified from 8 NEB10-Beta samples using the Monarch [®] Plasmid Miniprep Kit and digested with XmnI resulting in ≥90% conversion to a linearized fragment and no visible gDNA or RNA observed by gel electrophoresis in ≥75% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Nikki Samuelson
Production Scientist
11 Jan 2024



Julie Gaumont
Packaging Quality Control Inspector
07 Feb 2024